

UNITED STATES PATENT AND TRADEMARK OFFICE

UNITED STATES DEPARTMENT OF COMMERCE United States Patent and Trademark Office Address: COMMISSIONER FOR PATENTS P.O. Box 1450 Alexandria, Virginia 22313-1450 www.uspto.gov

ATTORNEY DOCKET NO. CONFIRMATION NO. FILING DATE FIRST NAMED INVENTOR APPLICATION NO. S63.2-10399 5949 10/063,125 03/22/2002 Lex P. Jansen **EXAMINER** 01/17/2006 WEBB, SARAH K VIDAS, ARRETT & STEINKRAUS, P.A. 6109 BLUE CIRCLE DRIVE ART UNIT PAPER NUMBER **SUITE 2000** MINNETONKA, MN 55343-9185 3731

DATE MAILED: 01/17/2006

Please find below and/or attached an Office communication concerning this application or proceeding.

Advisory Action

Application No.	Applicant(s)
10/063,125	JANSEN ET AL.
Examiner	Art Unit
Sarah K. Webb	3731

Before the Filing of an Appeal Brief --The MAILING DATE of this communication appears on the cover sheet with the correspondence address --THE REPLY FILED 12 December 2005 FAILS TO PLACE THIS APPLICATION IN CONDITION FOR ALLOWANCE. 1. 💢 The reply was filed after a final rejection, but prior to or on the same day as filing a Notice of Appeal. To avoid abandonment of this application, applicant must timely file one of the following replies: (1) an amendment, affidavit, or other evidence, which places the application in condition for allowance; (2) a Notice of Appeal (with appeal fee) in compliance with 37 CFR 41.31; or (3) a Request for Continued Examination (RCE) in compliance with 37 CFR 1.114. The reply must be filed within one of the following time periods: a) \square The period for reply expires $\underline{3}$ months from the mailing date of the final rejection. b) The period for reply expires on: (1) the mailing date of this Advisory Action, or (2) the date set forth in the final rejection, whichever is later. In no event, however, will the statutory period for reply expire later than SIX MONTHS from the mailing date of the final rejection. Examiner Note: If box 1 is checked, check either box (a) or (b). ONLY CHECK BOX (b) WHEN THE FIRST REPLY WAS FILED WITHIN TWO MONTHS OF THE FINAL REJECTION. See MPEP 706.07(f). Extensions of time may be obtained under 37 CFR 1.136(a). The date on which the petition under 37 CFR 1.136(a) and the appropriate extension fee have been filed is the date for purposes of determining the period of extension and the corresponding amount of the fee. The appropriate extension fee under 37 CFR 1.17(a) is calculated from: (1) the expiration date of the shortened statutory period for reply originally set in the final Office action; or (2) as set forth in (b) above, if checked. Any reply received by the Office later than three months after the mailing date of the final rejection, even if timely filed, may reduce any earned patent term adjustment. See 37 CFR 1.704(b). NOTICE OF APPEAL ___. A brief in compliance with 37 CFR 41.37 must be filed within two months of the date 2. The Notice of Appeal was filed on ___ of filing the Notice of Appeal (37 CFR 41.37(a)), or any extension thereof (37 CFR 41.37(e)), to avoid dismissal of the appeal. Since a Notice of Appeal has been filed, any reply must be filed within the time period set forth in 37 CFR 41.37(a). **AMENDMENTS** 3. The proposed amendment(s) filed after a final rejection, but prior to the date of filing a brief, will not be entered because (a) They raise new issues that would require further consideration and/or search (see NOTE below); (b) They raise the issue of new matter (see NOTE below); (c) They are not deemed to place the application in better form for appeal by materially reducing or simplifying the issues for appeal; and/or (d) They present additional claims without canceling a corresponding number of finally rejected claims. NOTE: (See 37 CFR 1.116 and 41.33(a)). 4. The amendments are not in compliance with 37 CFR 1.121. See attached Notice of Non-Compliant Amendment (PTOL-324). 5. Applicant's reply has overcome the following rejection(s): ___ 6. Newly proposed or amended claim(s) _____ would be allowable if submitted in a separate, timely filed amendment canceling the non-allowable claim(s). 7. 🛛 For purposes of appeal, the proposed amendment(s): a) 🖾 will not be entered, or b) 🗌 will be entered and an explanation of how the new or amended claims would be rejected is provided below or appended. The status of the claim(s) is (or will be) as follows: Claim(s) allowed: ___ Claim(s) objected to: Claim(s) rejected: 1-3,6-8,26-28 and 32-38. Claim(s) withdrawn from consideration: ____ AFFIDAVIT OR OTHER EVIDENCE 8. The affidavit or other evidence filed after a final action, but before or on the date of filing a Notice of Appeal will not be entered because applicant failed to provide a showing of good and sufficient reasons why the affidavit or other evidence is necessary and was not earlier presented. See 37 CFR 1.116(e). 9. The affidavit or other evidence filed after the date of filing a Notice of Appeal, but prior to the date of filing a brief, will not be entered because the affidavit or other evidence failed to overcome all rejections under appeal and/or appellant fails to provide a showing a good and sufficient reasons why it is necessary and was not earlier presented. See 37 CFR 41.33(d)(1). 10. The affidavit or other evidence is entered. An explanation of the status of the claims after entry is below or attached. REQUEST FOR RECONSIDERATION/OTHER 11. The request for reconsideration has been considered but does NOT place the application in condition for allowance because: See Continuation Sheet. 12. Note the attached Information Disclosure Statement(s). (PTO/SB/08 or PTO-1449) Paper No(s). 13. Other:

> Julian W. Wor JULIAN W. WOO

Continuation of 11. does NOT place the application in condition for allowance because: The arguments are not found to be persuasive. Applicant argues that the prior art does not provide teachings or suggestions to form a stent of a tungsten-rhenium alloy. Examiner agrees that the prior art combined in a 103 rejection should have a reasonable expectation of success. The office considers a reasonable expectation of success for forming a stent from a tungsten-rhenium alloy to be present in the prior art, as well as sufficient suggestion to form a stent from this alloy. Fontaine teaches that it is known to form stents from radiopaque metals, such as tantalum. Mayer teaches that tantalum, tungsten, and rhenium are suitable radiopaque material for use in stents. Evans teaches that a tungsten-rhenium alloy has good radiopacity and is suitable for use in medical devices. Bryskin simply provides known compositions for tungsten-rhenium alloys used for medical applications. There is sufficient evidence in the prior art to support a reasonable expectation of success for forming a stent from a tungsten-rhenium alloy.